



REPUBLIC OF KENYA

MINISTRY OF AGRICULTURE, LIVESTOCK, FISHERIES AND IRRIGATION

STATE DEPARTMENT FOR LIVESTOCK

ADDENDUM NO.1 TO THE OPEN NATIONAL TENDER

TENDER NO: MOALF&I/SDL/DLP/OT/03/2018/2019

TENDER FOR SUPPLY OF ULTRASONIC MILK ANALYZERS

ADDENDUM NO.1

Reference is made to the above tender which closes on 13th February, 2019 at 11:00am.

Supply and delivery of each and every Ultrasonic Milk Analyzer must be accompanied by a weighing scale. The specifications of the weighing device must at the minimum be as follows:

- | | |
|----------------------------|---|
| a) Type | - Hanging scale |
| b) Weight | - 2-3 kilograms |
| c) Unit of measurement | -kilograms, programmable to give output in litres |
| d) Capacity of measurement | - 0.1-100 kilograms |
| e) Display type | - Digital LCD display |
| f) Power source | - Rechargeable batteries |
| g) Connectivity | - Bluetooth enabled |
| h) Compatibility | - compatible with ultrasonic milk analyzers |
| i) Durability | - Durable and long lasting made of stainless steel |
| j) Accessories | - portable printer capable of combining milk quality and weight measurements in one print |

Take note that this Addendum No.1 will form part and parcel of the Tender document.

**HEAD, SUPPLY CHAIN MANAGEMENT SERVICES
FOR: PRINCIPAL SECRETARY**